

ALKALINE PHOSPHATASE (DST)

[DEA-p-NPP, Kinetic]

ENZOPAK

Last update 09-2020

Ref. CC2-ALK.002, 10x5 ml
CC2-ALK.02U, 10x10 ml

INTENDED USE

Reagent kit for quantitative estimation of Alkaline Phosphatase activity in serum or plasma.

PRODUCT HIGHLIGHTS

- Long Shelf Life
- High stability
- Scientific optimized formulation as per GSCC
- Pack sizes suitable for every laboratory
- Accuracy & reproducibility (precision) As per International Standard.
- Very well accepted by good customers

INTRODUCTION

Alkaline Phosphatase is a kinetic procedure based on the recommendations by German Society for Clinical Chemistry (GSCC). Most of the formulations by commercial manufacturers use substrate as *p*-NPP disodium hexahydrate, which deteriorates faster than *p*-NPP used in our formulation. The deterioration increases free *p*-NP, resulting in higher blanks. Our stabilised *p*-NPP resists deterioration and keeps the blank low all throughout the shelf life.


The concentration of the diethanolamine (DEA) buffer and substrate (*p*-NPP) in the reagent system estimate the phosphate transferase activity optimally, the DEA buffer system being a phosphate acceptor.

DIAGNOSTIC SIGNIFICANCE

Alkaline Phosphatase is present in high concentrations in liver, bone, placenta, intestine and certain tumors. Increase in Alkaline Phosphatase activity in serum or plasma is related to diseases of bone, biliary tract and liver. Decrease in activity is found in severe anemia, scurvy, kwashiorkor & cretinism.

PRINCIPLE

Alkaline Phosphatase in a sample hydrolyses para-nitrophenyl phosphate into para-nitrophenol and phosphate, in the presence of magnesium ions. The rate of increase in absorbance of the reaction mixture at 405 nm due to liberation of paranitrophenol is proportional to the alkaline phosphatase activity.



PRESENTATION

Store all reagents at 2-8°C

	Pack Size	
	10x5 ml	10x10 ml
• 1. Alk. Phosphatase (Substrate)	1 (10 Tablet)	1 (10 Tablet)
• 2. Alk. Phosphatase (Buffer)	1	2
• Reconstitution bottle	1	1

FINAL REAGENT COMPOSITION

Active Ingredients	Concentration
• <i>p</i> -NPP	20 mmol/L
• Sodium chloride	500 mmol/L
• Buffer	800mmol/L

pH 9.9± 0.5 at 25°C

Also contains non-reactive fillers and Stabilizers.

PRECAUTION

Alkaline Phosphatase is for *IN-VITRO* diagnostic use only. Reagent contains Sodium Azide, DO NOT INGEST.

PREPARATION OF WORKING REAGENT

For 10 x 5 ml

Carefully transfer one tablet of 1 alkaline phosphatase into the bottle containing 5 ml of 2 Alkaline Phosphatase (buffer). Mix gently to dissolve completely Wait for 15 minutes before use.

For 10 x 10 ml

Carefully transfer one tablet of 1 alkaline phosphatase into the bottle containing 10 ml of 2 Alkaline Phosphatase (buffer). Mix gently to dissolve completely Wait for 15 minutes before use.

REAGENT STORAGE AND STABILITY

Alkaline Phosphatase reagents are stable till the expiry stated on the label, when stored at 2-8°C.

The buffered substrate should be used on the same day. It may be stored in a refrigerator at 2-8°C for 30 days in a dark coloured bottle (provided in the kit). Para-nitrophenyl phosphate in solution slowly breaks down and paranitrophenol is liberated. If the blank absorbance of the buffered substrate exceeds 0.85 at 405 nm against distilled water, use options given under LINEARITY or discard the working reagent.

SPECIMEN COLLECTION

Fresh, clear serum, under fasting condition with no hemolysis is the specimen of choice.

Plasma collected using heparin as an anticoagulant may be used. Avoid anticoagulants like oxalate, citrate and EDTA.

REACTION PARAMETERS

- Type of Reaction : Kinetic/Increasing OD
- Wavelength : 405 nm
- Flowcell Temperature : 37°C
- Delay Time : 30 Seconds
- Interval : 30 Seconds
- No. of readings : 4
- Sample Volume : 20 µl (0.02 ml)
- Working Reagent Volume : 1.0 ml
- Factor : 2713
- Light Path : 1.0 cm
- Zero setting with : Distilled Water

TEST PROCEDURE

Pipette Into Test Tubes	TEST
Buffered Substrate (ml)	1.0
Sample (ml)	0.02

Mix and read absorbance at 30, 60, 90 and 120 seconds at 405 nm. Determine the mean change in absorbance per minute and calculate test results.

TEST RESULTS

Alkaline Phosphatase activity (IU/L) = $\Delta A/\text{min.} \times F$

Where F = 2713 (calculated on the basis of molar extinction coefficient for *p*-nitrophenol and ratio of total assay volume to sample volume).

NORMAL VALUES

CHILDREN :	250-770 IU/L	} At 37°C
(3-15) YEARS		
ADULTS :	100-250 IU/L	

ALKALINE PHOSPHATASE (DST)

[DEA-p-NPP, Kinetic]

ENZOPAK

Last update 09-2020

LINEARITY

Method is linear upto 1500 IU/L. For sample value exceeding the linearity limit, dilute samples suitably with 0.9% saline and repeat the assay. Apply dilution factor to calculate the test result.

For reconstituted reagent stored at 2-8°C, the following guidelines may be used for serum samples of very high enzyme activity.

- (a) If the initial absorbance is between 0.85 to 1.0 then the number of readings may be reduced to 2 with an interval of 30 seconds i.e. upto seconds only.
- (b) If the initial absorbance is between 1.0 to 1.2 then the number of readings may be reduced to 1 with an interval of 30 seconds i.e. upto seconds only.

NOTE

For laboratories using instruments with cuvette capacity more than 1.0 ml, sample and buffered substrate volumes should be increased proportionately.

REFERENCES

1. Recommendations of the German Society for Clinical Chemistry: Standardization of Methods for the estimation of Enzyme
2. Activity in Biological Fluids, J. Clinical Chemistry, Clinical Biochemistry 8. 192-192 (1972).



Regd. Office.3/7, B.I.D.C., Gorwa, Vadodara 390 016 (INDIA)

Web: www.reckondiagnosics.com Ph: +91-265-2281631

Email: mail@reckondiagnosics.com

An ISO 13485:2016 Certified Company

