ENZOPAK Last update 09-2020

INTENDED USE

Reagent kit for quantitative estimation of Gamma Glutamyl Transferase (transpeptidase) activity in serum or plasma.

INTRODUCTION

Gamma glutamyl transferase (GGT) was purified and characterized by Szewczuk and Baranowski. It catalyzes the transfer of the gamma glutamyl group from a gamma glutamyl peptide to an amino acid of another peptide.

In earlier methods, various substances such as glutathione, L-gamma-glutamyl-naphthylamide and L-gamma-amylanilide were used. Current methods use L-gamma-glutamyl-pnitroanilide (GPNA) as a substrate since end product is coloured and it permits a direct reaction rate measurement without deproteinization.

GGT is formulated on the Szasz method recommended by Scandinavian Society for Clinical Chemistry and Physiology. The method is kinetic, specific and sensitive.

DIAGNOSTIC SIGNIFICANCE

GGT is one of the most sensitive enzyme of the hepatobiliary system. Elevated serum gamma-GT levels are indicative of the disease of the liver, biliary tract and pancreas. In cases of metastatic carcinoma, viral hepatitis, chronic hepatitis, cholelithiasis, cholangitis and cholecystitis, gamma-GT levels are found to be elevated. Since GGT activity is not elevated in any bone disorders, the assay is considered as a valuable diagnostic aid for differentiation between bone and liver disease in conjuction with alkaline phosphatase.

PRINCIPLE

Gamma - glutamyl - p - nitroanilide (GPNA) and glycylglycine are converted by the action of GGT to p - nitroaniline and L gamma glutamyl glycylglycine. The rate of increase in absorbance at 405 nm due to the release of p-nitroaniline is directly proportional to the GGT activity.



PRESENTATION

All reagents to be stored at 2-8°C	No. of Bottles 20 x 1.1 ml
• 1 GGT (Substrate)	3 (7 Tablet)
• 2 GGT (Buffer)	2
 Reconstitution vial 	1

FINAL REAGENT COMPOSITION

Active Ingredients	Concentration
 Soluble GPNA 	≥3 mmol/L
• Buffer	100 mmol/L
 Glycyl glycine 	30 mmol/L
рН 8.5 <u>+</u> 0.1 at 25ºС	

Also contains non-reactive fillers and Stabilizers.

PRECAUTION

GGT is for *IN-VITRO* diagnostic use only. Reagent contains Sodium Azide. DO NOT INGEST.

PREPARATION OF WORKING REAGENT For 20 x 1.1 ml

Reconstitute one tablet of 1 GGT with 1.1 ml of 2-GGT. Mix gently to disolve the contents. Use after 5 minutes.



Regd. Office.3/7, B.I.D.C., Gorwa, Vadodara 390 016 (INDIA) Web: www.reckondiagnostics.com Ph: +91·265·2281631 Email: mail@reckondiagnostics.com An ISO 13485:2016 Certified Company

REAGENT STORAGE AND STABILITY

GGT reagent is stable until the expiry date printed on the label at $2\text{-}8^{\text{o}}\,\text{C}.$

Working reagent is stable for 14 days at $2-8^{\circ}$ C. Discard the working reagent if the absorbance reads 0.8 at 405 nm against distilled water.

SPECIMEN COLLECTION

Fresh, clear serum with no hemolysis is the specimen of choice. Plasma, collected with the use of heparin as an anticoagulant may also be used.

REACTION PARAMETERS

•	Type of Reaction	:	Kinetic/Increasing O.D.
٠	Wavelength	:	405 nm
٠	Flowcell Temperature	:	37°C / 30°C
•	Delay Time	:	60 Seconds
٠	Interval Time	:	30 Seconds
•	No. of reading	:	4
٠	Sample Volume	:	100 µl (0.1 ml)
٠	Factor	:	1158
•	Light Path	:	1.0 cm.
٠	Zero setting with	:	Distilled Water
	_		

TEST PROCEDURE

Pipette In To Test Tubes	TEST
Working Reagent (ml)	1.0
Sample (ml)	0.1

Mix immediately and read first absorbance of test exactly at 60 seconds and then, second, third and fourth at an interval of 30 seconds at 405 nm. Determine the mean change in absorbance per minute (ΔA /min) and use this for calculating test results.

TEST RESULTS

Serum GGT Activity (IU/L) =∆A/min X F

Where F = 1158 (Calculated on the basis of molar extinction coefficient for paranitroaniline and sample to total volume ratio.)

NORMAL VALUES

	30°C.	37°C.
Men (IU/L)	7-40	9-52
Women (IU/L)	4-25	5-32

LINEARITY

The method is linear upto 500 IU/L. For sample values higher than linear limit, dilute the sample suitably with 0.9 % saline and repeat the assay. Apply dilution factor to calculate test results.

NOTE

For laboratories using instruments with cuvette capacity more or less than 1.0 ml. sample volume and working reagent volume may be proportionately increased or decreased.

REFERENCES

 SZASZ G. Reaction rate method for gamma-glutamyl transferase activity in serum. Clin. Chem. 22.2051-2055 (1976).

