

URIC ACID-L (R1-R2)

(Uricase, Enzymatic)

ENZOPAK

Last update 04-2023

Ref. CC3-URK.20M, 2x20 ml
CC3-URK.20MU, 6x20 ml

INTENDED USE

Reagent kit for quantitative estimation of Uric acid in serum or plasma.

INTRODUCTION

Uric acid is produced by the action of xanthine oxidase on xanthine and hypoxanthine, which are products of nucleic acid degradation. The chemical method of estimation of uric acid is based on the uric acid forming a blue phosphotungstate complex. However, this method is not specific as other reducing substances are present in serum or plasma which continues to produce colour by phosphotungstate complex. In 1980, Fossati et, al described a modified Trinder's method for the assay of uric acid.

Uric Acid is formulated on this method with the advantage that the method is specific and sensitive.


DIAGNOSTIC SIGNIFICANCE

Uric Acid levels in any subject may vary during a day from time to time and from day to day. The changes represent the contents in diet and metabolism besides pathological condition.

The diagnostic values showing increased in levels are found in kidney failure and gout. Elevated levels are found in acute infectious diseases, severe uremia, toxemia of pregnancy, leukemia and some malignant diseases or conditions

PRINCIPLE

Uricase is a very specific enzyme acting on Uric Acid, end products being allantoin and hydrogen peroxide. Peroxidase is used to utilize hydrogen peroxide (proportional to uric acid concentration) to convert chromogens to coloured complex. The intensity of colour produced is proportional to uric acid concentration and is measured photometrically at 546 nm (530-570) nm or with YELLOW filter).



PRESENTATION

All reagents to be stored at 2-8°C	No. of Bottles	
	2x20 ml	6x20 ml
• 1 Uric Acid – Liquid	2x16 ml	6x16 ml
• 2 Uric Acid – Liquid	2x4 ml	6x4 ml
• Uric-Acid Standard (6 mg/dL)	1x2.5 ml	1x2.5 ml

FINAL REAGENT COMPOSITION

ACTIVE REAGENTS	Concentration
• Buffer	100 mmol/L
• 4-AAP	1 mmol/L
• POD	≥ 2000 U/L
• Buffer	100 mmol/L
• TOOS	0.5 mmol/L
• Uricase	≥ 600 U/L

pH 7.0 ± 0.1 at 25°C

Uric Acid Standard (6 mg/dl)

Also contains non-reactive fillers & stabilizers.

PRECAUTION

Uric acid is for *IN-VITRO* diagnostic use only.
Reagent Contains Sodium Azide, DO NOT INGEST.

PREPARATION OF WORKING REAGENT

Carefully transfer the contents of 1 bottle of 2 Uric Acid into the bottle of 1 Uric Acid. Mix well. Wait for 2 minutes before use.

Alternatively for flexibility as much of working reagent may be made as and when desired by mixing 4 parts of 1 Uric Acid liquid and 1 part of 2 Uric Acid liquid.

REAGENT STORAGE & STABILITY

All reagents included in the kit are stable at 2-8°C until the expiry date indicated on the label.

Working reagent is stable for 4 weeks at 2-8°C. A slight rise in the blank reading due to stored life will not affect the test results.

SPECIMEN COLLECTION

Fresh, clear fasting serum with no hemolysis is the specimen of choice. However, plasma collected from blood using heparin as an anticoagulant may be used.

REACTION PARAMETERS

- Type of Reaction : End Point
- Wavelength : 546 nm (530-570 nm)
- Flowcell Temperature : 37°C
- Incubation : 10 min. at 37°C
- Std. Concentration : 6 mg/dL
- Sample Volume : 50 µl (0.05 ml)
- Reagent Volume : 1.0 ml
- Light Path : 1.0 cm
- Zero setting with : Reagent Blank

TEST PROCEDURE

Pipette into Test tubes	Blank	Standard	Test
Working Reagent (ml)	1.0	1.0	1.0
Standard (ml)	-	0.05	-
Sample (ml)	-	-	0.05

Mix and incubate at 37°C for 10 min. and read absorbance of test and standard against reagent blank at 546 nm (530-570 nm or yellow filter).

STABILITY OF FINAL REACTION MIXTURE

The colour of reaction mixture is stable for one hour when protected from direct light.

TEST RESULTS

$$\text{Uric Acid Concentration (mg/dl)} = \frac{\text{Absorbance of Test}}{\text{Absorbance of Standard}} \times 6$$

To convert Uric Acid concentration (mg/dl) to micromoles/liter uses the following equation,
Micromoles/lit. = mg/dl x 59.5

LIMITATIONS FOR INTERFERENCE

As per studies carried out for interference. Following results were obtained.

- No Interference from Hemoglobin upto 500 mg/dl.
- No Interference from free Bilirubin upto 15 mg/dl.
- No Interference from Lipemic (Measured as Triglycerides) upto 1000 mg/dl.

URIC ACID-L (R1-R2)

(Uricase, Enzymatic)

ENZOPAK

Last update 04-2023

NORMAL VALUES

(SERUM)

Male : 3.2 - 7.0 mg/dl (190.4 - 416.5 μ mol/L)

Female : 2.6 - 6.0 mg/dl (154.7 - 357 μ mol/L)

(URINE)

Urine Uric acid content: 400-900 mg/24 hours.

LINEARITY

This method is linear upto 20 mg/dl. For sample values higher than linearity limit, dilute the samples suitably with 0.9 % saline and repeat the assay. Apply proper dilution factor to calculate the final test results.

REFERENCES

1. TRIVEDI, R.C., REBBAR, L.BERKA, E., AND Strong, I: Clin.Chem 24:1908(1978).
2. FOSSATI, P., PRINCIPIE, L., BERTI, G:Clin Chem 26, 227-231 (1980).



Regd. Office.3/7, B.I.D.C., Gorwa, Vadodara 390 016 (INDIA)

Web: www.reckondiagnosics.com Ph: +91-265-2281631

Email: mail@reckondiagnosics.com

An ISO 13485:2016 & GMP Certified Company

