

ALKALINE PHOSPHATASE

(DEA-p-NPP, Kinetic)

ENZOPAK

Last update 09-2020

Ref. CC1-ALK.02M, 50x1.1 ml

INTENDED USE

Reagent kit for quantitative estimation of Alkaline Phosphatase activity in serum or plasma.

PRODUCT HIGHLIGHTS

- Long shelf life
- High stability (Tableted Substrate)
- Optimised formulation as per GSCC
- Convenient Pack sizes
- Accuracy & reproducibility (precision) of International Standard

INTRODUCTION

Alkaline Phosphatase is a kinetic procedure based on the recommendation by German Society for Clinical Chemistry (GSCC).

Most of the formulations by commercial manufacturers use substrate as *p*-NPP disodium hexahydrate salt which deteriorates faster than *p*-NPP used in our formulation. The deterioration increases free *p*-NP, resulting in higher blanks. Our stabilised *p*-NPP resists deterioration and keeps the blank low all throughout the shelf life.

The concentrations of the diethanolamine (DEA) buffer and substrate (*p*-NPP) in the reagent system estimate the phosphate transferase activity optimally, the DEA buffer system being a phosphate acceptor.

DIAGNOSTIC SIGNIFICANCE


Alkaline Phosphatase is present in high concentrations in the liver, bone, placenta, intestine and certain tumors.

Increase in Alkaline Phosphatase activity in serum or plasma is related to diseases of bone, biliary tract and liver.

Decrease in activity is found in severe anemia, scurvy, kwashiorkar & cretinism.

PRINCIPLE

Alkaline Phosphatase in a sample, hydrolyses para-nitrophenyl phosphate into paranitrophenol and phosphate, in the presence of magnesium ions. The rate of increase in absorbance of the reaction mixture at 405 nm due to liberation of paranitrophenol is proportional to the alkaline phosphatase activity.



PRESENTATION

All reagents to be stored at 2-8°C.	No. of Bottle/Blister 50 x 1.1 ml
• 1 Alkaline Phosphatase (Substrate)	5 (10 Tablets)
• 2 Alkaline Phosphatase (Buffer)	1
• Reconstitution vial	2

FINAL REAGENT COMPOSITION

Active Ingredients	Concentration
• <i>p</i> -NPP	16.3 mmol/L
• Sodium chloride	1000 mmol/L
• Buffer	1000 mmol/L

pH 9.9± 0.1 at 25°C

Also contains non-reactive fillers and Stabilizers.

PRECAUTION

Alk. Phosphatase is for *IN-VITRO* diagnostic use only.

Reagent contains Sodium Azide, DO NOT INGEST.

PREPARATION OF WORKING REAGENT

Dissolve one tablet (1 ALK. PHOSPHATASE) in 1.1 ml of buffer (2 ALK. PHOSPHATASE) to make buffered substrate. For quick dissolution crush the tablet prior to addition of buffer. Keep for 15 minutes before use.

REAGENT STORAGE AND STABILITY

Alkaline Phosphatase reagents are stable till the expiry stated on the label, when stored at 2-8°C.

The buffered substrate should be used on the same day. It may be stored in a refrigerator (2-8°C) for 10 days in a dark coloured bottle. Para-nitrophenyl phosphate in solution slowly breaks down and para-nitrophenol is liberated. If the blank absorbance of the buffered substrate exceeds 0.85 at 405 nm against distilled water, use options given under LINEARITY or discard the working reagent.

SPECIMEN COLLECTION

Fresh, clear serum, under fasting condition with no hemolysis is the specimen of choice.

Plasma collected using heparin as an anticoagulant may be used. Avoid anticoagulants like oxalate, citrate and EDTA.

REACTION PARAMETERS

- Type of Reaction : Kinetic/Increasing OD
- Wavelength : 405 nm
- Flowcell Temp. : 37°C
- Delay Time : 30 Seconds
- Interval : 30 Seconds
- No. of reading : 4
- Sample volume : 20 microliters (0.02 ml)
- Working Reagent Volume : 1.0 ml
- Factor : 2713
- Light Path : 1.0 cm.
- Zero setting with : Reagent Blank

TEST PROCEDURE

Pipette into Test Tubes	TEST
Buffered substrate (ml)	1.0
Sample (ml)	0.02

Mix and read absorbance at 30, 60, 90 and 120 seconds at 405 nm. Determine the mean change in absorbance per minute and calculate test results.

TEST RESULTS

Alkaline Phosphatase activity (IU/L) = $\Delta A / \text{min} \times F$

Where F = 2713 [calculated on the basis of molar extinction coefficient for *p*-nitrophenol and ratio of total assay volume to sample volume].

NORMAL VALUES

	At 37°C
Children (3-15 years)	: 250-770 IU/L
Adults	: 100-250 IU/L

ALKALINE PHOSPHATASE

(DEA-p-NPP, Kinetic)

ENZOPAK

Last update 09-2020

LINEARITY

Method is linear upto 1500 IU/L. For sample value exceeding the linearity limit, dilute samples suitably with 0.9 % saline and repeat the assay. Apply dilution factor to calculate the test results.

For reconstituted reagent stored for more than 3 days at 2-8°C the following guidelines may be used for serum samples of very high enzyme activity:

- a) If the initial absorbance is between 0.85 to 1.0 then the number of readings may be reduced to 2 with an interval of 30 seconds i.e. upto 90 seconds only.
- b) If the initial absorbance is between 1.0 to 1.2 then the number of readings may be reduced to 1 with an interval of 30 seconds i.e. upto 60 seconds only.

NOTE

For laboratories using instruments with cuvette capacity more than 1.0 ml, sample and buffered substrate volumes should be increased proportionally.

REFERENCES

1. Recommendations of the German Society for Clinical Chemistry
2. Standardization of Methods for the estimation of Enzyme Activity in Biological Fluids, J. Clinical Chemistry, Clinical Biochemistry 8. 182 -192 (1972)



Regd. Office.3/7, B.I.D.C., Gorwa, Vadodara 390 016 (INDIA)

Web: www.reckondiagnosics.com Ph: +91-265-2281631

Email: mail@reckondiagnosics.com

An ISO 13485:2016 Certified Company

