

CREATININE-L (Single Liquid)

(JAFFE MODIFIED)

ENZOPAK

Last update 09-2020

Ref. CC3-CRL.08N, 5x25 ml
CC3-CRL.08NU, 5x60 ml

INTENDED USE1

Reagent kit for quantitative estimation of creatinine in serum or urine.

INTRODUCTION

Creatinine is an end product of creatinine metabolism which is a part of energy utilization mechanism in muscles. The quantity of creatinine generated depends marginally on the dietary protein intake. The serum creatinine level depends directly on the body muscle mass and is affected by intense muscular stress.

In 1886, Jaffe developed a method for the assay of creatinine based upon the reaction between creatinine and sodium picrate. In 1904, Folin employed this reaction for the quantitative estimation of creatinine, in urine. Since then, a number of modifications have been formulated to reduce interference from proteins and non-creatinine substances like keto-acids, glucose, ascorbic acid etc.

Creatinine liquid is a modified formulation of Jaffe's reaction which does not require deproteinization. The picrate and alkali reagents are formulated in such a way measuring the rate of reaction, interference from non-creatinine substances present in the sample is almost eliminated.

DIAGNOSTIC SIGNIFICANCE

Creatinine is used to assess renal function, however, serum creatinine levels do not start to rise until renal function has decreased by at least 50%. Congestive heart failure, shocks and mechanical obstruction of urinary tract may also contribute to an elevated level of serum creatinine. An elevated serum creatinine level due to obstruction may rapidly fall when the obstruction is removed by surgery.

Many times serum urea/creatinine ratio is used for assessment of kidney function and differential diagnosis. Creatinine clearance test is carried out only for assessment of kidney function.

PRINCIPLE

Creatinine present in the serum or urine reacts with alkaline picrate to form a colored complex. The rate of formation of colored complex is directly proportional to creatinine concentration. This rate of reaction (intensity of color produced) is measured photometrically at 510 nm and is compared with that of the standard.

Creatinine + Alkaline Picrate → Creatinine Picrate Complex

PRESENTATION

All reagents to be stored at 2-8°C.

| | No. of Bottles/ Vials | |
|-------------------------------------|-----------------------|---------|
| | 5x25 ml | 5x60 ml |
| • CREATININE REAGENT (Ready to Use) | 5 | 5 |
| • CREATININE STANDARD (2 mg/dl) | 1 | 1 |

FINAL REAGENT COMPOSITION

| Active Ingredients | Concentration |
|--------------------|---------------|
| • Picric Acid | 10 mmol/L |
| • Sodium Hydroxide | 100 mmol/L |

Creatinine Standard (2 mg/dl)

Also contains non-reactive fillers and Stabilizers.

PRECAUTION

Creatinine single reagent is for *IN-VITRO* diagnostic use only. Avoid contact of reagents with skin, eyes and clothes. Use automated pipetting devices.

REAGENT STORAGE AND STABILITY

All reagents included in the kit are stable at 2-8°C until the expiry date stated on the label.

SPECIMEN COLLECTION

Fresh, clear serum with no hemolysis is the specimen of choice. Plasma prepared using heparin as an anticoagulant may also be used.

REACTION PARAMETERS

- Type of Reaction : Fixed Time/Two Point / Initial Rate
- Wavelength : 510 nm
- Flowcell/cuvette Temp. : 30°C
- Delay Time : 20 seconds
- Interval : 60 seconds
- Measuring Time : 80 seconds
- Standard/Sample Volume : 100 µl (0.1ml)
- Standard Concentration : 2 mg/dl
- Reagent Volume : 1.0 ml
- Light Path : 1.0 cm
- Zero setting with : Distilled Water

Note: For instruments using cuvette capacity of 2.5 ml, use sample and standard volume 200 microliters (0.2 ml) and reagent 2.5 ml.

TEST PROCEDURE

| PIPETTE INTO TEST TUBES | STANDARD (ST) | TEST (TS) |
|-------------------------|---------------|-----------|
| REAGENT (ml) | 1.0 | 1.0 |
| STANDARD (ml) | 0.1 | - |
| SAMPLE (ml) | - | 0.1 |

Mix and aspirate. Record the absorbance of Standard (ST) and Test (TS) at 20 seconds (ST₁, TS₁) and again at 80 seconds (ST₂, TS₂) at 510 nm, against distilled water.

TEST RESULTS

$$\text{Creatinine Concentration} = \frac{(TS_2 - TS_1)}{(ST_2 - ST_1)} \times 2$$

To convert (mg/dl) to micromol/liter, use the following equation;

$$\text{micromol/liter} = (\text{mg/dl}) \times 88.5$$

LIMITATIONS FOR INTERFERENCE

As per studies carried out for Interference, following results were obtained.

- No Interference from Hemoglobin upto 50 mg/dl.
- No Interference from free Bilirubin upto 12 mg/dl.
- No Interference from Lipemic (Measured as Triglycerides) upto 500 mg/dl. = Sample volume in ml

NORMAL VALUES: Serum Creatinine

MEN : 0.9 - 1.4 mg/dl

WOMEN : 0.6 - 1.2 mg/dl

CREATININE-L (Single Liquid)

(JAFTE MODIFIED)

ENZOPAK

Last update 09-2020

CREATININE ESTIMATION IN URINE

For Creatinine estimation in urine, dilute the sample suitably with distilled water and follow the procedure to calculate test results by applying dilution factor. A dilution of 1:15 or 1:100 is suggested.

Creatinine concentration in urine (gms/lit)

$$= \frac{(TS2-TS1)}{(ST2-ST1)} \times 2 \times \frac{\text{Dilution factor}}{100}$$

NORMAL VALUES : Urine Creatinine

MEN : 1.0 – 2.0gms/24 hours

WOMEN : 0.8 – 1.5gms/24 hours

LINEARITY

This method is linear upto 12 mg/dl. For sample values higher than linearity limit, dilute the samples suitably with 0.9% saline and repeat the assay. Apply dilution factor to obtain test results.

REFERENCES

KAPLANA., SZABO, L.L., Clinical Chemistry : Interpretation and Techniques, Lea & Febiger, Philadelphia (1983).



Regd. Office.3/7, B.I.D.C., Gorwa, Vadodara 390 016 (INDIA)

Web: www.reckondiagnosics.com Ph: +91-265-2281631

Email: mail@reckondiagnosics.com

An ISO 13485:2016 Certified Company

